

ALTERNATIVE - Von Frey Test ALTCP_VFR_001

Parameters and Metadata

ALTERNATIVE - Baseline: tabulation ALTCP_VFR_001_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Baseline: final score ALTCP_VFR_002_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Baseline: average final score ALTCP_VFR_003_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_002_001',1)

ALTERNATIVE - Baseline: PWT force ALTCP_VFR_004_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Baseline: average PWT force ALTCP_VFR_00

5_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_004_001',1)

ALTERNATIVE - Test 1: tabulation ALTCP_VFR_006_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 1: final score ALTCP_VFR_007_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 1: average final score ALTCP_VFR_008_01 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_007_001',1)

ALTERNATIVE - Test 1: PWT force ALTCP_VFR_009_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 1: average PWT force ALTCP_VFR_010_0

01 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_009_001',1)

ALTERNATIVE - Test 2: tabulation ALTCP_VFR_011_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 2: final score ALTCP_VFR_012_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 2: average final score ALTCP_VFR_013_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_012_001',1)

ALTERNATIVE - Test 2: PWT force ALTCP_VFR_014_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Increments: Minimum 1

ALTERNATIVE - Test 2: average PWT force ALTCP_VFR_015_0

01 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Derivation: meanOfIncrements('TCP_VFR_014_001',1)

ALTERNATIVE - Challenge ALTCP_VFR_016_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: CFA,

ALTERNATIVE - Site of challenge injection ALTCP_VFR_017_0

01 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Plantar surface of right hindpaw,

ALTERNATIVE - General anaesthetic for challenge injection ALTCP_VFR_018_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Isoflurane,

ALTERNATIVE - Number of runs per test ALTCP_VFR_019_001 |

v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: 2, 1,

ALTERNATIVE - Number of trials per run ALTCP_VFR_020_001 |

v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: 5,

ALTERNATIVE - Minimum interval between filament presentation ALTCP_VFR_021_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: min

Description: ALTERNATIVE -

Options: 2,

ALTERNATIVE - Number of repeats with same filament ALT

TCP_VFR_022_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Once, 2-3, if response not obvious in first 2 then 3rd performed,

ALTERNATIVE - Minimum acclimatisation period ALTTCP_VF

R_023_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE -

Options: 1,

ALTERNATIVE - Paws tested ALTTCP_VFR_024_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Right hindpaw,

ALTERNATIVE - Time between baseline measurement and challenge

ALTTCP_VFR_025_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE -

Options: 1, 2,

ALTERNATIVE - Time between challenge and test 1

ALTTCP_VFR_026_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE -

Options: 22, 24,

ALTERNATIVE - Time between challenge and test 2

ALTTCP_VFR_027_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE -

Options: 142, 144,

ALTERNATIVE - Tetrad manufacturer ALTCP_VFR_028_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: IITC,

ALTERNATIVE - Tetrad dimensions ALTCP_VFR_029_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: cm

Description: ALTERNATIVE -

Options: 12.5 cm H x 10 cm W x 10 cm L,

ALTERNATIVE - Tetrad material ALTCP_VFR_030_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Plexiglass,

ALTERNATIVE - Tetrad colour-opacity ALTCP_VFR_031_001 | v1.0

0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Clear,

ALTERNATIVE - Inset material ALTCP_VFR_032_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Plexiglass,

ALTERNATIVE - Inset colour-opacity ALTCP_VFR_033_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Clear, White,

ALTERNATIVE - Grid material ALTCP_VFR_034_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Black painted metal,

ALTERNATIVE - Grid hole size ALTCP_VFR_035_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: mm

Description: ALTERNATIVE -

Options: Hexagonal 8 mm corner to corner,

ALTERNATIVE - Filament set manufacturer ALTCP_VFR_036_0

01 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Aesthesio, Stoetling,

ALTERNATIVE - Filament set model ALTCP_VFR_037_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Aesthesio Precise Tactile Sensory Evaluator 20 piece kit,

ALTERNATIVE - Filament material ALTCP_VFR_038_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: Nylon,

ALTERNATIVE - Range of filaments used (target force) ALT

TCP_VFR_039_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Options: 0.02 - 1.4,

ALTERNATIVE - Starting filament (target force) ALTCP_VFR_

040_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: g

Description: ALTERNATIVE -

Options: 0.16,

ALTERNATIVE - Starting filament (filament number) ALTTCP

_VFR_041_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: 5,

ALTERNATIVE - Date filaments last calibrated ALTTCP_VFR_04

2_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

ALTERNATIVE - Experimenter ID ALTTCP_VFR_043_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

ALTERNATIVE - Disinfectant ALTCP_VFR_044_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Description: ALTERNATIVE -

Options: Clidox/Ethanol, Quaternary Ammonia (Coverage Plus),

ALTERNATIVE - Scaling parameter X ALTCP_VFR_045_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

ALTERNATIVE - Scaling parameter B ALTCP_VFR_046_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

ALTERNATIVE - Testing methodology ALTTCP_VFR_047_001 | v1.

0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE -

Options: SUDO,
