

ALTERNATIVE - Hargreaves Test ALTUCD_H RG_001

Parameters and Metadata

ALTERNATIVE - Radiant heat light source manufacturer AL

TUCD_HRG_032_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: IITC,

ALTERNATIVE - Site of challenge injection ALTUCD_HRG_013_0

01 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Plantar surface of right hind paw,

ALTERNATIVE - Testing chamber dimensions ALTUCD_HRG_0

29_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: cm

Description: ALTERNATIVE - null

Options: 12.7 cm H x 10.16 cm W x 10.16 cm L, 12.5 cm H x 10 cm W x 10 cm L,

ALTERNATIVE - Disinfectant ALTUCD_HRG_035_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Coverage Plus Working Solution, Clidox and 70% Ethanol, 1% Virkon,

ALTERNATIVE - Test 2: average latency to withdrawal ALTUCD_HRG_006_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE -

Derivation: meanOfIncrements('UCD_HRG_005_001',1)

ALTERNATIVE - Time between challenge and test 2 ALTUCD_HRG_017_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE - null

Options: 144, 48,

ALTERNATIVE - Experimenter ID ALTUCD_HRG_034_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

ALTERNATIVE - Test 1: average latency to withdrawal ALTUCD_HRG_004_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE -

Derivation: meanOfIncrements('UCD_HRG_003_001',1)

ALTERNATIVE - Anaesthetic for challenge injection ALTUCD

_HRG_014_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Isoflurane,

ALTERNATIVE - Lid colour-opacity ALTUCD_HRG_028_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Clear,

ALTERNATIVE - Glass base heated during acclimatisation

ALTUCD_HRG_008_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Yes,

ALTERNATIVE - Baseline: average latency to withdrawal A

LTUCD_HRG_002_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE -

Derivation: meanOfIncrements('UCD_HRG_001_001',1)

ALTERNATIVE - Length of acclimatisation period ALTUCD_H

RG_007_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: min

Description: ALTERNATIVE - null

Options: 60, 30, 30 - 60,

ALTERNATIVE - Test intensity ALTUCD_HRG_021_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: %

Description: ALTERNATIVE - null

Options: 100, 25, 30, 15,

ALTERNATIVE - Glass base manufacturer ALTUCD_HRG_030_001

| v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: IITC,

ALTERNATIVE - Idle intensity ALTUCD_HRG_020_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: %

Description: ALTERNATIVE - null

Options: 2, 3, 10,

ALTERNATIVE - Tetrad colour-opacity [ALTUCD_HRG_024_001](#) | v1

.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Clear,

ALTERNATIVE - Tetrad manufacturer [ALTUCD_HRG_022_001](#) | v1.

0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: IITC,

ALTERNATIVE - Time between challenge and test 1 ALTUCD

_HRG_016_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE - null

Options: 24,

ALTERNATIVE - Minimum time between stimulus presentation ALTUCD_HRG_011_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: min

Description: ALTERNATIVE - null

Options: 2,

ALTERNATIVE - Tetrad material ALTUCD_HRG_023_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Acrylic,

ALTERNATIVE - Baseline: latency to withdrawal ALTUCD_HR

G_001_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE - null

Increments: 1, 2, 3,

ALTERNATIVE - Glass base temperature: settings ALTUCD_H

RG_018_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: C

Description: ALTERNATIVE - null

Options: 34, 30, 32,

ALTERNATIVE - Inset colour-opacity ALTUCD_HRG_026_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: White opaque, Clear,

ALTERNATIVE - Radiant heat light source model ALTUCD_HRG_033_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: 390 Plantar test head (Item #300TH),

ALTERNATIVE - Time between baseline measurement and challenge injection ALTUCD_HRG_015_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: Hours

Description: ALTERNATIVE - null

Options: 0.5-1, 2, 24,

ALTERNATIVE - Challenge ALTUCD_HRG_012_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: CFA,

ALTERNATIVE - Test 2: latency to withdrawal ALTUCD_HRG_0

05_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE - null

Increments: 1, 2, 3,

ALTERNATIVE - Inset material ALTUCD_HRG_025_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Acrylic,

ALTERNATIVE - Lid material ALTUCD_HRG_027_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Acrylic,

ALTERNATIVE - Test 1: latency to withdrawal ALTUCD_HRG_03_001 | v1.0

seriesParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE - null

Increments: 2, 1, 3,

ALTERNATIVE - Glass base model ALTUCD_HRG_031_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Model 400 Heated Base (Item #400G),

ALTERNATIVE - Number of repeats per paw ALTUCD_HRG_010_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Description: ALTERNATIVE - null

Options: Right hind paw tested 3 times,

ALTERNATIVE - Maximum exposure time ALTUCD_HRG_009_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Unit Measured: s

Description: ALTERNATIVE - null

Options: 30, 20,

ALTERNATIVE - Glass base temperature: measured ALTUCD

_HRG_019_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Unit Measured: C

Description: ALTERNATIVE - null

Options: 29-32,
