

Clinical Chemistry JAXLA_CBC_003

Purpose

Clinical chemistry determines biochemical parameters in plasma including enzymatic activity, specific substrates and electrolytes.

Ontological description: MP:0001545 – blood physiology abnormalities.

Experimental Design

- **Minimum number of animals** : 7M + 7F
- **Age at test**: Week 78
- **Sex**: We would expect the results of this test to show sexual dimorphism

Equipment

1. Clinical chemistry analyser
2. Vortex
3. Refrigerated centrifuge
4. Eppendorf tubes
5. Pipettes (200-1000 ul)

Procedure

Set up the clinical chemistry analyser and perform QC analyses of the control reagents in accordance with the equipment guidelines.

Sample collection and preparation:

- a. Collect the appropriate volume of blood required (160-200l of plasma), for the clinical chemistry analyser being used for assessment, in gel tube containing lithium Heparin with the relevant blood collection procedure (see IMPC protocol Blood collection by retro-orbital puncture). Time of day for collection is in the morning, starting no earlier than 07:30.
- b. Keep whole blood samples in a bag on wet ice until centrifugation. Centrifuge for 10 minutes at 5000 x g in a refrigerated centrifuge set at 8°C. If plasma samples cannot be analysed immediately, keep them in the fridge until analysis.
- c. Analysis of samples is optimally done on the day of collection. When not possible the plasma samples can be stored at 2-8°C. If samples require storage for > 48 hours,

freeze plasma at -20 °C in single aliquots. All samples are allowed to come to room temperature prior to analysis.

- d. Use plasma samples undiluted or diluted to a ratio of 1:2 with deionised water if the volume is insufficient.
- e. Plasma samples that were frozen or stored in the fridge should be vortexed briefly and centrifuged again at ~5000 x g for 2-3 minutes immediately prior to analysis. If necessary, remove fibrin clots using a wooden applicator.

Analysis:

Samples that produce results that lie outside the linear range for a specific assay have to be re-tested. In some cases it may be necessary to dilute samples with water to bring test results into range.

Notes

Blood collection for Clinical Chemistry and Hematology is usually performed as a non-fasting, terminal procedure but can be performed as a non-terminal procedure under certain circumstances. Mice from the terminal procedure may be used for subsequent gross pathology and other procedures included in terminal assessments. Whole blood (for Hematology) and plasma (for Clinical Chemistry) require different collection tubes so two independent samples are required from each mouse.

The information about the date of the experiment, that is the date when the measurement is performed, is an important parameter which is to be submitted in the Experiment xml file (dateOfExperiment="2013-02-28").

Dilution. Dilution of blood is highly discouraged, but is allowed when the total necessary amount is not obtained. If dilution is necessary then the assays should be done in one run.

Hemolysis. Two fields currently exist to capture metadata information about the hemolysis status in the clinical chemistry plasma samples. The first is the LIH Hemolysis severity score which can only be performed by clinics who run one of the Beckman Coulter AU-series of analysers. Such clinics are encouraged to capture and submit the hemolysis score of the LIH test in this field. Clinics who do not have an AU analyser are encouraged to use the second /alternative field which is simply titled Hemolysis. Simply enter "slight", "moderate", or "marked" based on whether the sample is visibly haemolysed or not. Provision of this information is not compulsory and it is suggested that any clinic completes at least one field or the other (not both).

Data QC

1. Plasma samples must be free of Fibrin clots in order to be analysed.
2. Badly haemolysed samples should be discarded.
3. Each morning, all parameters are tested with control sera (see ESLIM_015_001_Annex_3: Controls for biochemistry on AU400). Some parameters are tested with control serum level 1 (Beckman Coulter System Reagent, ODC0003)

and control serum level 2 (Beckman Coulter System Reagent, ODC0004), which consists of lyophilised human plasma with a normal and a pathological concentration. Other parameters are tested with specific controls from other suppliers.

4. Controls are thawed and vortexed before utilisation and loaded according to the analyser's display. Control values must lie within the acceptable range indicated by the manufacturer, otherwise the specific tests must be recalibrated and specific measurements repeated. Controls can be stored in 200l aliquots at -20°C for up to 1 week.

Metadata and examples

Metadata	Example
Equipment ID	ID of the machine used when more than 1 is used having same model and manufacturer. E.g. machine 1, machine 2, machine Minnie, machine Mickey Mouse, etc.
Equipment manufacturer	Manufacturer of the equipment. E.g. Olympus Diagnostics.
Equipment model	Model of the equipment. E.g. AU400
Blood collection tubes	The tubes used for blood collection. E.g. Sarstedt Li-Heparin gel tubes or Kabe Labortechnik Lithium heparin coated tubes.
Anaesthesia used for blood collection	The drug used for anaesthesia during blood collection. E.g. Isoflurane.
Method of blood collection	Concise description of the method used for blood collection. E.g. retro-orbital puncture.
Anticoagulant	Anticoagulant drug used for blood collection. E.g. Li-Heparin.
Samples kept on ice between collection and analysis	Yes/No.

Storage temperature from blood collection till measurement	E.g. 2°C
Sample status	Indicate if the sample were frozen (analysis on the same day of collection not possible) or fresh (analysis on the same day of collection). E.g Fresh/Frozen.
Plasma dilution	Dilution is highly discouraged but if necessary indicate here. E.g. “No dilution” or 1:2. Note that results submitted to DCC are assumed to be already corrected for any dilutions made.
ID of blood collection SOP	ID of the protocol followed for blood collection. Can be a center specific protocol. E.g. ESLIM_024_001.
Date and time of blood collection	Time of day for collection is in the morning, starting no earlier than 07:30. E.g. Year, month, day, time.
Date of measurement	The day of blood analysis. Year, month, day.
Hemolysis status	If no AU analyser score is provided, indicate here the gauged degree of hemolysis. E.g. slight/moderate/marked.
Blood collection experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who collected the blood. E.g. Harw_001, or 1/2/3.
Blood analysis experimenter ID	An ID of any format to be used coherently both inside the same procedure and for all procedures indicating the experimenter who analyzed the blood. E.g. Harw_001, or 1/2/3.
Date equipment last calibrated	Most recent date in which the equipment (or any part of) used in the procedure was subject to a calibration event.
Date and time of sacrifice	The date and time when the mouse is sacrificed.

Parameters and Metadata

Transferrin JAXLA_CBC_031_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: mg/dl

Fructosamine JAXLA_CBC_020_001 | v1.2

simpleParameter

Req. Analysis: false Req. Upload: false Is Annotated: true

Unit Measured: umol/l

Blood collection experimenter ID JAXLA_CBC_049_001 | v1.1

procedureMetadata

Req. Analysis: false Req. Upload: true Is Annotated: false

Creatine kinase JAXLA_CBC_028_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: U/l

Alanine aminotransferase JAXLA_CBC_013_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: U/l

Phosphorus JAXLA_CBC_010_001 | v1.6

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Anesthesia used for blood collection JAXLA_CBC_036_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Injection narcosis with Sodium Pentobarbital (Somnopentyl),
Injection narcosis with Ketamine (100mg/kg)/ Xylazine (10mg/kg)/Antipamezole (Antisedan, 1mg/kg),
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg),
Injection narcosis with Ketamine (100mg/kg)/Xylazine (10mg/kg),
Injection narcosis with Sodium Pentobarbital (Pentobarb, 0.1ml), No,
Injection narcosis with Ketamine (137mg/kg)/Xylazine (6.6mg/kg),
Gas anaesthesia with Isoflurane,
Injection narcosis with Ketamine (110mg/kg)/Xylazine (11mg/kg)/ Antipamezole (Antisedan, 1mg/kg),
Injection narcosis with Sodium Pentobarbital (Euthatal),
Injection narcosis with Tribromoethanol (Avertin),

C-reactive protein JAXLA_CBC_032_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/l

Aspartate aminotransferase JAXLA_CBC_012_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: U/l

Lipase JAXLA_CBC_021_001 | v1.1

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: U/l

Alpha-amylase JAXLA_CBC_023_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: U/l

Cholesterol ratio JAXLA_CBC_058_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Derivation: div('JAXLA_CBC_015_001', 'JAXLA_CBC_016_001')

Blood collection tubes JAXLA_CBC_039_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Options: Kabe Labortechnik Lithium heparin coated tubes,
BD Microtainer Lithium Heparin/PST Gel Blood Tube, Eppendorf 1.7ml,
BD Microtainer Lithium Heparin Tube, Kabe Labortechnik 1000ul Lithium Heparin,
Sarstedt Li-Heparin gel tubes, TERUMO CAPIJECT Lithium heparin coated tubes,
Greiner MiniCollect Lithium Heparin 1ml,

Magnesium JAXLA_CBC_054_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

Albumin JAXLA_CBC_007_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: g/l

Sample type JAXLA_CBC_056_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Plasma, Serum,

Chloride JAXLA_CBC_003_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

ID of blood collection SOP JAXLA_CBC_045_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: sop.inv.063, PHENO_CBC, RIKENMPP_004a_003, ESLIM_024_001, sop.inv.019,

Sample dilution JAXLA_CBC_044_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: 1:4, 1:2, Neat serum, 1:5, Yes (by Equipment, automatically), Varies, 1:3, Neat plasma,

Date and time of sacrifice JAXLA_CBC_040_001 | v1.1

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Hemolysis status JAXLA_CBC_048_001 | v1.1

procedureMetadata

Req. Analysis: false **Req. Upload:** false **Is Annotated:** false

Options: None, Slight, Moderate, Marked,

Method of blood collection JAXLA_CBC_037_001 | v1.0

procedureMetadata

Req. Analysis: true **Req. Upload:** true **Is Annotated:** false

Options: Heart puncture, Jugular vein, Cardiac puncture, Retro-orbital puncture, Tail vein,

Total protein JAXLA_CBC_006_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: g/l

Total cholesterol JAXLA_CBC_015_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Date and time of blood collection JAXLA_CBC_046_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Difficult bleed JAXLA_CBC_055_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Options: Yes, No,

Reagent manufacturer JAXLA_CBC_059_001 | v1.0

procedureMetadata

Req. Analysis: true **Req. Upload:** false **Is Annotated:** false

Options: Microgenics, Beckman Coulter, Wako and Sekisui,

Glycerol JAXLA_CBC_027_001 | v1.4

simpleParameter

Req. Analysis: false **Req. Upload:** false **Is Annotated:** true

Unit Measured: mmol/l

Ferritin JAXLA_CBC_030_001 | v1.3

simpleParameter

Req. Analysis: false **Req. Upload:** false **Is Annotated:** true

Unit Measured: ng/ml

Iron JAXLA_CBC_011_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

Thyroxine JAXLA_CBC_053_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: ug/dl

HDL-cholesterol JAXLA_CBC_016_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Sodium JAXLA_CBC_001_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

Lactate dehydrogenase JAXLA_CBC_022_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: U/l

Anticoagulant JAXLA_CBC_038_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Sodium Heparin, Heparine, Lithium Heparin, No,

Glycosilated hemoglobin A1c (HbA1c) JAXLA_CBC_052_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: %

Blood analysis experimenter ID JAXLA_CBC_051_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Free fatty acids JAXLA_CBC_026_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

Fasting JAXLA_CBC_057_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Sixteen hours before bleeding, No, Four hours before bleeding,

Triglycerides JAXLA_CBC_017_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Equipment manufacturer JAXLA_CBC_034_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Roche, JEOL (Siemens), Cobas, Beckman Coulter, Hitachi, Olympus Diagnostics,

Total bilirubin JAXLA_CBC_008_001 | v1.4

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Date equipment last calibrated JAXLA_CBC_050_001 | v1.2

procedureMetadata

Req. Analysis: false

Req. Upload: false

Is Annotated: false

LIH (Hemolysis Severity - available on AU analysers) JAXLA

_CBC_019_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: false

Storage temperature from blood collection till measurement JAXLA_CBC_041_001 | v1.3

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Unit Measured: C

Options: 18-22, -80, 4, 2,

Equipment model JAXLA_CBC_035_001 | v1.0

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: Hitachi 917, Integra 400 Plus, AU 400, UniCel 600 Pro, AU 480, JCA-BM6070, JCA-BM2250 (Advia 2400), 7020, AU 680, DxC AU 700, Cobas,

Creatinine JAXLA_CBC_005_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Samples kept on ice between collection and analysis JAXLA _CBC_042_001 | v1.1

procedureMetadata

Req. Analysis: true

Req. Upload: true

Is Annotated: false

Options: No, Yes,

Glucose JAXLA_CBC_018_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

LDL-cholesterol JAXLA_CBC_025_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mg/dl

Calcium JAXLA_CBC_009_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

Equipment ID JAXLA_CBC_033_001 | v1.0

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Alkaline phosphatase JAXLA_CBC_014_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: U/l

Potassium JAXLA_CBC_002_001 | v1.3

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: mmol/l

Uric acid JAXLA_CBC_029_001 | v1.2

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: umol/l

Urea (Blood Urea Nitrogen - BUN) JAXLA_CBC_004_001 | v1.5

simpleParameter

Req. Analysis: false

Req. Upload: true

Is Annotated: true

Unit Measured: mg/dl

UIBC (unsaturated iron binding capacity) JAXLA_CBC_024_001 | v1.0

simpleParameter

Req. Analysis: false

Req. Upload: false

Is Annotated: true

Unit Measured: umol/l

Sample status JAXLA_CBC_043_001 | v1.1

procedureMetadata

Req. Analysis: false

Req. Upload: true

Is Annotated: false

Options: Frozen, Fresh, Fresh and frozen,
