OCT HMGULA_OCT_001

Purpose

To detect abnormalities in eye morphology.

Experimental Design

- **Minimum number of animals :** 7M + 7F
- **Age at test:** Week 77
- **Sex:** We do not expect the results of this test to show sexual dimorphism

Procedure

1. Turn on the OCT and start the database
2. Anaesthetize mouse
3. Prepare mouse eyes with drops and place contact lens (focal length 10 mm) on the right eye
4. Enter mouse data in the “Create new patient file” area and switch to the “ Acquisition” window
5. Move the OCT camera to the right position and activate measurement modus
6. Place mouse collaterally to the OCT camera on the right side of a platform that is fixed in front of the OCT lens
7. Search the contact lens in the live picture of the fundus image field and place the pupil of the mouse eye in the centre of the window
8. Move the OCT camera such that OCT lens and contact lens come close to each other
9. Focus the fundus picture by slightly moving up/down or forward/backward, and lock the camera at this position
10. Take *en face* fundus images
11. Set the "Ref.Arm“ ruler such that the section of the retina is placed in the centre of the blue rectangle
12. Set the mode of measurement on SD-OCT modus
13. Move the green horizontal line in the fundus image field to the optic nerve level
14. Save images of retinal sections

15. Move the OCT camera to the left position
16. Repeat measurement procedure for the left eye

Notes

- As a minimum, all abnormalities should be imaged.
  - Where capacity permits, all mice can be imaged
- Majority of parameters can be analysed using the standard approach for assessing categorical data. To increase power for analysis purposes, where an abnormality is detected in the left, right or both eyes, the data may be combined to generate one "abnormal" category.

Data QC

Image QC is typically performed during data collection to ensure high quality images are captured whilst eyes are dilated etc.

Parameters and Metadata

Fundus retina HMGULA_OCT_001_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

OCT description HMGULA_OCT_002_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** false **Is Annotated:** false

---

Left fundus number of main vessels HMGULA_OCT_003_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Right fundus number of main vessels HMGULA_OCT_004_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Left retinal thickness HMGULA_OCT_005_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** true

Unit Measured: µm

Right retinal thickness HMGULA_OCT_006_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** true

Unit Measured: µm

Left fundus pigmentation HMGULA_OCT_007_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Right fundus pigmentation HMGULA_OCT_008_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Left optic disc HMGULA_OCT_009_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Right optic disc HMGULA_OCT_010_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Left retinal layers HMGULA_OCT_011_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Right retinal layers HMGULA_OCT_012_001 | v1.0

simpleParameter

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Datetime of measurement HMGULA_OCT_013_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

---

Equipment manufacturer HMGULA_OCT_014_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: Heidelberg Engineering,

Mouse status HMGULA_OCT_015_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: Anaesthetized,

Topical Agents HMGULA_OCT_016_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: Atropine,

Equipment model HMGULA_OCT_017_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: Spectralis,

Equipment ID HMGULA_OCT_018_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** false **Is Annotated:** false

---

Date equipment last calibrated HMGULA_OCT_019_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** false **Is Annotated:** false

---

Dilation Method HMGULA_OCT_020_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: 0.5 % Atropine Solution, Atropine,

General Anesthetic HMGULA_OCT_021_001 | v1.0

procedureMetadata

Req. Analysis: false **Req. Upload:** true **Is Annotated:** false

Options: Ketamine+Xylazine, Ketamin 0.1 mg/g, Xylazin 0.01mg/g bodyweight,

---